Experiment HM-4: Stimulus Response, Work, Summation, and Tetanus in Human Muscle

Equipment Required

PC or Mac Computer

IXTA, USB cable, Power supply for IXTA

HV stimulator lead wires for IXTA

Disposable stimulating electrodes

SMT-220 Striated muscle transducer

Surgical tape or Velcro^{tm} straps

Thread

Weights (10, 20, 30, 40, 50 grams)

IXTA Isolated Stimulator

The IXTA has a high voltage stimulus isolator designed to deliver constant current to the nerve or muscle being studied. In situations where the resistance (R) along the path of the current increases, the voltage (V) increases to maintain the current (I in V = IR, Ohm's Law). The ability of the IXTA to adjust the voltage to deliver the required current is known as voltage compliance. The upper limit of this compliance by the IXTA is set at 100 Volts.

Constant current devices differ from constant voltage devices when presented with an increase in resistance, like the dehydration of the conductive gel under the electrodes. As pointed out earlier, a constant current stimulator is voltage compliant. In constant voltage stimulators, the current delivered to the tissue decreases as the resistance increases because the power supply of the constant voltage device is not designed to deliver additional current.

Although the IXTA can generate up to 100 Volts, the current delivered by the unit is limited to a maximum of 20 milliamperes, for a maximum duration of 10 milliseconds per pulse, and a maximum frequency of 50 pulses per second (Hz). At these levels, the maximum amount of power delivered by the IXTA will not cause injury or tissue damage.

The current is selected using the Stimulator Control Panel. The HV Stimulator can deliver a maximum output of twenty milliamperes

The duration, frequency, and number of stimulus pulses generated by the stimulator are also controlled by making changes to the values in the Stimulator Control Panel. The initial values of the pulses generated by the IXTA are programmed by the same settings file that configured the recording software. For example, if a pulse from the IXTA is programmed for a duration of 1 millisecond and a frequency of 1 Hz, the stimulator will generate a stimulus pulse with the same duration and frequency.

IWorx Sample Lab

IXTA Stimulator Setup

1. Place the IXTA on the bench near the subject.

Warning: Before connecting the IXTA stimulating electrodes to the subject, check the Stimulator Control Panel to make sure the amplitude value is set to zero (0).

Note: Disconnect the subject from the IXTA prior to powering off the device.

Figure HM-4-S1: The IXTA stimulating electrodes.

Figure HM-4-S2: The front panel of the IXTA with the stimulating electrodes connected correctly.

2. Instruct the subject to remove all jewelry before beginning the experiment.
3. Connect the color-coded stimulator lead wires to the High Voltage Current Stimulator. Make sure you push the safety connector of each lead wire into the appropriate socket as far as possible.

Positioning of the Stimulating Electrodes

Before attaching the SMT-220 striated muscle transducer to the side of the subject's hand, determine the positions for the stimulating electrode that cause the muscles of the little finger to create a strong twitch response.

iWorx Home Lab

1. Use an alcohol swab to clean and scrub two areas on the back of the subject's hand. One area is in the center of the hand, and the second area is along the lateral edge of the hand between the base of the little finger and the wrist.
2. Obtain a couple of disposable electrodes. Attach one electrode to the center of the back of the hand. The electrode should be half way between the first knuckle of the middle finger and the wrist. This electrode is the positive (red) stimulating electrode.
3. Attach the second electrode about one centimeter above the lateral edge of the hand, half way between the first knuckle of the little finger and the wrist. This electrode is the negative (black) stimulating electrode.
4. Snap the ends of the H.V. Output lead wires onto the stimulating electrodes, so that:
    - the red (+) lead is snapped on the electrode in the center of the back of the hand,
    - the black (-) lead is snapped on the electrode at the lateral edge of the hand.



Figure HM-4-S3: A. Placement of the stimulating electrodes on the back of the right hand. B. Placement of weight (10) across the inside of the hand and sensor on the outer edge of little finger.

5. Click on the Stimulator button in the LabScribe toolbar to open the Stimulator control panel on the LabScribe Main window. Compare the values set for the IXTA stimulator to the values programmed on the Preferences Dialog window, which can be viewed by selecting Preferences from the Edit menu on the LabScribe Main window. Make any changes to the parameters that are necessary and click on the **Apply** button to complete the changes. The **Apply** button will change to **Fire** when recording.

Warning: Make sure the Amplitude on the Stimulator Control Panel is set to zero.

6. Click Record and then click **Fire** on the stimulator control panel. There should be no response from the subject's finger since the current output is zero. Continue to record.

iWorx Student Lab

Note: *As the current output is increased in the next couple of steps, ask the subject to indicate when he or she first feels any tingling under the negative stimulating electrode. Minor movement of the subject's finger can usually be seen when the stimulus current is raised another milliamp or two.*

7. Enter the amplitude value into the Stimulator Control Panel and click APPLY which then changes to FIRE – click FIRE to send out the stimulation. For each "1" amplitude increase it is equal to a current output of 2 milliamperes (mA). Remind the subject to indicate the first occurrence of tingling. If no finger movement is seen, the stimulus current is below the threshold current of the most sensitive muscle fibers controlling the finger's movement. Threshold is the current level that is needed to create a muscle fiber contraction.

Note: *The amplitude required to cause a finger twitch will differ between subjects. Some subjects require as low as 2.5 milliamperes of current (1.25 amplitude) to create a maximum response, while other subjects may require 7 or more milliamperes (3.5 or more amplitude) to create the strongest response.*

8. If a finger twitch does not occur at 2 mA, change the Amplitude by an additional 0.5. Click the Apply button. Check for tingling and finger movement. Increase the current output in increments of 0.5 until the subject's finger twitches with the largest range of motion.

NOTE – the pinkie may not twitch and another finger will twitch – just move the twitch sensor to the finger that is actively twitching. It does not need to be the pinkie.

9. If the stimulus current has been raised to 10 in the Amplitude box and the subject feels tingling but no finger movement is observed, click Stop and adjust the position of the negative stimulating electrode before making additional increases in the stimulus current.
    - Make sure to stop the recording before moving the electrodes.
    - Move the negative stimulating electrode a centimeter closer to the positive stimulating electrode.
    - Stimulate the subject's hand. Raise the stimulus current as high as 5 mA. If the response from the subject's hand is miniscule, or non-existent, place the negative electrode at different points along an imaginary line on the back of the hand, from the base of the little finger to the lateral edge of the wrist.
10. Find the lowest stimulus current that creates the largest twitch from the subject's finger, click Stop to turn off the stimulator.
11. Set the current at this level for Exercise 1.

Note: *The lowest stimulus that causes the largest possible response is called the maximal stimulus. Any stimulus that is above the maximal level is known as supra-maximal. Currents above threshold and below maximal are called sub-maximal.*

IWorX Sample Lab

SMT-220 Setup

1. Insert the connector of the SMT-220 striated muscle transducer into the “Temp” input on the front of the IXTA. The sensor in the SMT-220 is a flexible resistive element that generates a signal that is proportional to range of flexion of the sensor and the finger to which it is attached.
2. The SMT-220 generates a signal only when flexed in one direction. Locate the label on the sensor that indicates the side of the sensor that will need to face outwardly from the finger.
3. Align the SMT-220 along the side of the little finger so the outside of the sensor is oriented away from the finger. Use a Velcro^{tm} strap, or piece of surgical tape, to attach the tip of the sensor to the end of the little finger. Use a longer Velcro^{tm} strap, or piece of surgical tape, to attach the base of the sensor along the lateral edge of the palm of the hand.

Figure HM-4-S4: The SMT-220 striated muscle transducer with a label indicating the side of the sensor that should face outwardly from the finger.

Figure HM-4-S5: The IXTA with the stimulating electrodes and the SMT 220 twitch sensor connected correctly.

iWorx Lab

Experiment HM-4: Stimulus Response, Work, Summation, and Tetanus in Human Muscle

WARNING - The Stimulator should only be used for the method of application for which the Stimulator is intended as shown in the directions below.

Note: Disconnect the subject from the IXTA prior to powering off the device.
If the subject feels a painful stimuli, then the electrodes are in the wrong place.

NOTE: If using the IXTA and built in HV stimulator – all changes in Amplitude are entered directly into the Stimulator Control Panel. Click "APPLY" to make any changes. "APPLY" will change to the word "FIRE". When recording, click "FIRE" to fire the actual stimulus.

Exercise 1: Properties of a Finger Twitch

Aim: To become familiar with using a high voltage stimulator and to record a finger twitch with and without a weight attached to the finger.

Approximate Time: 15 minutes

Procedure

1. Instruct the subject to relax and place the hand that is used for the experiment on the bench, with its palm up.
2. Click the Stimulator Preferences icon on the LabScribe toolbar to open the stimulator control panel if it does not open automatically.



Figure HM-4-L1: The LabScribe toolbar.



Figure HM-4-L2: The stimulator control panel.

Sample Lab

3. Check the values for the stimulus parameters that are listed in the stimulator control panel on the Main window: the pulse amplitude (Amp) should be set to 5.0V or mA, the pulse width (W) to 5 ms, the frequency (F) to 1 Hz; and, the number of pulses (#pulses) to 15. The value for a stimulus parameter can be changed by typing the value in the window next to the label of the parameter. Click the Apply button to finalize the changes.
4. When using the IXTA change the amplitude in the software to set the stimulus current to the level determined in the Setup section of this lab.

Remember to click APPLY to change the stimulus parameters and FIRE to generate the pulses.

5. Type the **< Subject's Name>** in the Mark box. Click Record. Click the mark button to mark the recording. Continue to record for 15 to 20 seconds. Click the Stop button.
6. Select Save As in the File menu, type a name for the file. Click on the Save button to save the data file.
7. Leave all the stimulus parameters at the same settings.

Data Analysis

1. Scroll through the recording of the finger twitches. Notice each finger twitch has two phases, a quick contraction phase followed by a slow relaxation phase. Since the duration of the stimulus pulses are short and their frequency is low, the muscles of the finger have the opportunity to fully relax between twitches as indicated by the return of the twitch amplitude to the same baseline between twitches.
2. Use the Display Time icons to adjust the Display Time of the Main window to show at least three complete twitches on the Main window.
3. Data can be collected from the Main window or the Analysis window. If you choose to use the Analysis window, click on the Analysis window icon in the toolbar.
4. The mathematical functions, V2-V1 and T2-T1 should appear on screen. Values for V2-V1 and T2-T1 on each channel are seen in the table across the top margin of each channel, or to the right of each graph.
5. The functions in the channel pull-down menus of the Analysis window can also be used to enter the names and values of the parameters from the recording to the Journal. To use these functions:
    - Place the cursors at the locations used to measure the amplitude and times of each finger twitch.
    - Transfer the names of the mathematical functions used to determine the amplitude and times to the Journal using the Add Title to Journal function in the Finger Twitch Channel pull-down menu.

iworx Sample Lab

6. Once the cursors are placed in the correct positions for determining the amplitude and times of each finger twitch, the values of the parameters can be recorded in the on-line notebook of LabScribe by typing their names and values directly into the Journal, or on a separate data table.
7. Transfer the values for the amplitude and period to the Journal using the Add Ch. Data to Journal function in the Finger Twitch Channel pull-down menu.



Figure HM-4-L3: Finger twitches recorded without a weight attached to the finger, seen on the upper channel. The signal of the high voltage stimulus isolator is seen on the lower channel at a frequency of 1Hz.

Any movement, voluntary or involuntary, will be recorded. It is important that the hand is fully relaxed and in a comfortable position.

8. On the Finger Twitch channel, use the mouse to click on and drag the cursors to specific points on the recording to measure the following parameters:
    - Twitch Amplitude. To measure the amplitude of the twitch, place one cursor on the baseline at the point just before the twitch starts to develop, and the second cursor on the peak of the twitch. On the Finger Twitch channel, the value for V2-V1 is the amplitude of that twitch.
    - Contraction Time, which is the time it takes the amplitude of the twitch to rise to its peak. To measure the contraction time of the twitch, keep the cursors in the same positions used to measure the twitch amplitude. On the Finger Twitch channel, the value for T2-T1 is the contraction time of that twitch.

IWorx Sy

- Relaxation Time, which is the time it takes the amplitude of the twitch to return to baseline. To measure the relaxation time of the twitch, place one cursor on the peak of the twitch, and the second cursor at the point where the amplitude returns to the baseline. On the Finger Twitch channel, the value for T2-T1 is the relaxation time of that twitch.



Figure HM-4-L4: Finger twitches stimulated by maximal current at a frequency of 1 Hz.

Question

How does the contraction time of a typical finger twitch compare to the relaxation time?

Exercise 2: The Effect of Weight on Finger Twitches

Aim: To determine the amount of work performed by the finger with different weights, or loads, attached to the finger.

Approximate Time: 30 minutes

Procedure

1. Check the stimulus parameters set in the Stimulator control panel on the LabScribe Main window. These settings should be the same as the ones used in Exercise 1.
2. Set the stimulus current to the maximal level as used in Exercise 1.

iWorx Sa

3. Attach a thread to the last joint of the finger being used in the experiment. The thread should be long enough to go across the ring, middle, and index fingers of the same hand, and also allow the weight to hang over the edge of the table. Attach a weight to the other end of the thread. The weight could be a pen, a pencil, or washers.
4. Have the subject move their hand to the edge of the table, so the palm is up and the thread with the weight is hanging along the side of the table.
5. Type the weight of the object, that is on the thread, in the Mark box. Click Record. Click the mark button to mark the recording. Continue to record for 15 seconds. Click the Stop button.
6. Add another 10 gram weight to end of the thread and repeat Step 5.
7. Continue to add weight to the thread in 10 gram increments. Repeat Step 5 each time you add weight to the thread until the amplitude of the finger twitch is undetectable.
8. Select Save in the File menu on the LabScribe window.
9. Remove the weights and thread from the subject's finger.

Data Analysis

1. Scroll to the beginning of the recording for this exercise. Locate the section of data in which finger twitches appeared when the stimulus threshold was reached. Click AutoScale to maximize the size of the twitch on the window.
2. Use the Display Time icons to adjust the Display Time of the Main window to show a segment with four or five twitches on the Main window.
3. Use the same techniques used in Exercise 1 to measure and record the amplitudes (V2-V1) of three twitches recorded at each weight.
4. For each weight, determine the average amplitude of three finger twitches.
5. Use the average twitch amplitude at each weight to calculate the work performed on that weight. For this exercise, use the following equation:

Work = Weight (g) x Average Amplitude of Twitch (mv) with that Weight

6. In Excel plot the work performed at each weight as a function of the weight attached to the finger.

Questions

1. How do the amplitudes of the finger twitches with different weights compare to each other?
2. Why did the amount of work decrease when heavier weights were used?
3. Is there any correlation between the weight moved and the contraction or relaxation times of the twitches?

IWorx Sample Lab



Figure HM-4-L5: The relationship between work performed at each weight and the weight moved by a finger twitch

Exercise 3: Stimulus-Response

Aim: To examine the effect of stimulus current on the amplitude of finger movement.

Approximate Time: 30 minutes

Procedure

1. Remind the subject to relax and place the hand that is used for the experiment on the bench, with its palm up.
2. Check the stimulus parameters set in the stimulator control panel. The duration of the stimulus pulses should be 5 milliseconds (ms), their frequency should be 1 Hz, and the number of pulses (#p) should be 0 (zero). Setting the number of pulses to zero makes the stimulator fire continuously. If the values of these parameters were changed, click the **Apply** button on the Stimulator control panel to effect the change
3. Make sure the Amplitude is set to 0. Type **Zero** in the Mark box. Click Record and click the mark button to mark the recording.
4. While recording, type **0.5 mA** in the Mark box. Increase the current output by 0.5mA. Click **APPLY and then FIRE.** Click the mark button to mark the recording. Record at this stimulus current for ten to fifteen seconds.
5. Repeat Step 4, in increments of 0.5 mA, until the stimulus current reaches the maximal level. Click the Stop button.
6. Open the File menu and select Save to save the data file.

iWorx Sample Lab

Data Analysis

1. Scroll to the beginning of the data for this exercise. Locate the segment of the finger twitches that occurred at threshold stimulus current. Click AutoScale to maximize the size of the twitch on the window. For any stimulus currents that did not cause finger movement, report the amplitude of the finger twitch as zero (0).
2. Use the Display Time icons to adjust the Display Time of the Main window to show a segment with four or five twitches on the Main window.
3. Use the same techniques used in Exercise 1 to measure and record the amplitudes (V2-V1) of three twitches recorded for each stimulus current that causes a response.
4. For each stimulus current, determine the average amplitude of three finger twitches.
5. In Excel plot the average amplitude of the finger twitch at each stimulus current as a function of the stimulus current.



Figure HM-4-L6: Average finger twitch amplitude as a function of the stimulus current

Questions

1. At which stimulus current did the finger first show a twitch? Why didn't a finger twitch occur at a lower stimulus current?
2. How does the finger twitch amplitude increase when the stimulus current is increased?
3. Is there a stimulus current at which the amplitude of the finger twitch is maximized?
4. Above a certain stimulus current, the amplitude of the finger twitch no longer increases. Why does this happen?

IWorX Lab

Exercise 4: Summation and Tetanus

Aim: To monitor the contraction and relaxation of the finger twitch in relation to the stimulus frequency.

Approximate Time: 30 minutes

Procedure

1. Remind the subject to relax and place the hand that is used for the experiment on the bench, with the palm up.
2. Check the stimulus parameters set in the stimulator control panel. The duration of the stimulus pulses should be 5 milliseconds (ms), their frequency should be 1 Hz, and the number of pulses should be set to 40.
3. During the course of this exercise, the stimulus frequency will be changed to demonstrate summation and tetanus. After any stimulus parameter is changed, click the **Apply** button on the Stimulator control panel to make the change. Click **Fire** to fire the stimulation.
4. Set the stimulus current to the maximal level as used in Exercise 1.
5. Type **1 Hz** in the Mark box. Click Record and click the mark button. Do not stop the recording There should be forty twitches on the recording.
6. Change the stimulus frequency to 2 Hz. Click the **Apply** button to put any frequency change into effect.
7. Repeat Step 5 with the stimulus frequency set to 2. Type **2** in the Mark box. When the subject is ready – click **Fire**.
8. Repeat Step 7 for stimulus frequencies of 3, 4, 5, and then 10, 15, and 20 Hz. Remember to click the **Apply** to make the changes and **Fire** to send the stimulation.
9. When performing the experiment at the highest frequency (20 Hz) keep recording after complete tetanus is reached. Continue to record for at least 10 seconds as the tension in the muscles of the finger begins to relax.
10. When the muscles in the finger are fully relaxed, click Stop to halt the recording.
11. Select Save in the File menu.

Data Analysis

1. Scroll to the beginning of the data for this exercise. Locate the segment of the finger twitches that occurred at a frequency of 1 Hz. Click AutoScale to maximize the size of the twitches on the window.
2. Use the down arrow to the right of the Mark button to locate the mark for 1 Hz, click that mark and the recording will bring you right to that section of data.
3. If needed, use the Display Time icons to adjust the Display Time of the Main window to show the twitches at that frequency on the Main window.

IWorx Sample Lab



Figure HM-4-L7: Amplitudes of finger twitches at 10, 15, and 20 Hz from left to right. Recordings from 10 and 15 Hz show incomplete tetanus. Complete tetanus occurs at 20 Hz.

4. Use the same techniques used in Exercise 1 to measure and record the amplitudes (V2-V1) of two adjacent twitches recorded at a stimulus frequency of 1 Hz.
5. Move to the segment of data recorded at the next stimulus frequency by clicking the down arrow next to the mark button and clicking on the next mark. If the twitches at this frequency are all about the same amplitude, measure and record the amplitudes of two adjacent twitches.
6. Repeat Step 5, until the recording of the finger twitches show summation. **Summation** occurs when the time between twitches is not long enough to allow the muscle to completely relax to its baseline level of tension. Another twitch or contraction, following in quick succession to the first, will add on the first twitch at the current level of tension for the first twitch. The third twitch will add on to the current tension of the second twitch, and so on. Since the twitches were maximal, this summation is a form known as **mechanical summation**.
7. At stimulus frequencies where summation occurred, measure and record the amplitude of the first twitch in the series and the amplitude of the tallest twitch in the series.
8. If a level of relatively constant tension occurs during contractions at high frequency, the phenomenon is called tetanus. If small relaxations are still detectable along the level of constant tension, a state of incomplete tetanus exists (5 Hz, 10 and 15 Hz). If there are no detectable relaxations along the region of constant tension, a state of complete tetanus exists (20 Hz).

Work Sam



Figure HM-4-L8: Summation of finger twitches occurring at a stimulus frequency of 5Hz.

Questions

1. At which frequency did summation first appear? At which frequencies does incomplete tetanus occur? At which frequency did complete tetanus first occur?
2. How much greater than the amplitude of a single twitch is the amplitude of the tallest twitch during summation or incomplete tetanus?
3. How much greater than the amplitude of a single twitch is the amplitude of complete tetanus?
4. After the stimulator is turned off, the relaxation time after tetanus is much longer than relaxation time after a single twitch. Why?
5. What is the name for the decrease in the level of constant tension that occurs during tetanus?

IWorx Sample